

Tulsa Gas Technologies, Inc.

4809 South 101st East Ave.
Tulsa, OK 74146
Phone 918-665-2641 Fax 918-665-2657

Calibration Report

DATE: September 5, 2019
Test # 1
Test Date: 9/4/2019
Tested By: Dylan Higginbotham
Test Type Acceptance

Test Site:

Tulsa Gas Technologies, Inc.
4809 South 101st East Ave.
Tulsa, OK 74146

CONTACT: Jerry Kelley
PHONE: 918 665 2641

Work Order Number:

AMTEK072419

Customer Name:

Tulsa Gas Technologies
4809 South 101st East Ave.
Tulsa, OK. 74146

Dispenser Model Number:

T-7203-8CNG50W

Dispenser Serial Number:

19081265

Eprom Number:

1.04.13

	SCALE	DISPENSER	PERCENT of ERROR
HOSE 1	MASS/GGE	GGE	
	10.56	10.600	-0.38%
	10.84	10.810	0.28%
	10.53	10.510	0.19%
HOSE 2	9.84	9.850	-0.10%
	10.30	10.330	-0.29%
	9.79	9.780	0.10%

EXPECTABLE TOLERANCE IS + - 2% FOR EACH BATCH
AND + - .8% BETWEEN BATCHES PER EACH HOSE

HOSE 1

Prover Starting Mass	0.0000								
Prover 1st Batch	10.5600								
Dispenser 1st Batch	10.600	-0.38 %							
Prover 2nd Batch	21.4000								
Dispenser 2nd Batch	10.810	0.28 %							
Prover 3rd Batch	31.9300								
Dispenser 3rd Batch	10.510	0.19 %							
Prover 4th Batch	0.000								
Dispenser 4th Batch	0.000	100.00 %							
STARTING PROVER READING									
Lb/GGE	0.0000	SCALE	10.5600	SCALE/GGE	10.560	GALLONS	10.600	PERCENT of ERROR	NEW GGE FACTOR
1	10.5600	10.5600	10.560	10.560	10.600	10.600	-0.38	1.00	
	21.4000	10.8400	10.840	10.840	10.810	10.810	0.28		
	31.9300	10.5300	10.530	10.530	10.510	10.510	0.19		
									0.03%

HOSE 2

Prover Starting Mass	0.0000								
Prover 1st Batch	9.8400								
Dispenser 1st Batch	9.850	-0.10 %							
Prover 2nd Batch	20.1400								
Dispenser 2nd Batch	10.330	-0.29 %							
Prover 3rd Batch	29.9300								
Dispenser 3rd Batch	9.780	0.10 %							
Prover 4th Batch									
Dispenser 4th Batch		100.00 %							
STARTING PROVER READING									
Lb/GGE	0.0000	SCALE	9.8400	SCALE/GGE	9.840	GALLONS	9.850	PERCENT of ERROR	NEW GGE FACTOR
1	9.8400	9.8400	9.840	9.840	9.850	9.850	-0.10	1.00	
	20.1400	10.3000	10.300	10.300	10.330	10.330	-0.29		
	29.9300	9.7900	9.790	9.790	9.780	9.780	0.10		
									-0.10%