

Product Code

Serial ID

Order ID

Line


Item

Customer Tag

26011539

CNG050S290NCAAEZZZ

15623624



5616237

1.1

2

Process

Detail

Process ID : 9.31175738

Process Time : 2025.06.05 9:57:47

Process Stand : SSF1E@SSCCL:1

Stand Uncertainty : +/-0.030%

Fluid : H2O

100% Rate : 38.6 KG/MIN

Pickoff : 1

Max Rate P/T : 41.78 PSIG/21.4 C

Results

Status : PASS

D1 : 0

D2 : 1

K1 : 4060.75

K2 : 4228.09

DT : 4.25

FD : 0

DTG : 0

DFQ1 : 0

DFQ2 : 0

FlowCal : 141.524.50

FFQ : 0

FTG : 0

DensCal : 04061042284.25

FCF : 141.52

FT : 1.5

By: *Bed Carter*
RAZVAN, TATAR

Technician

This certificate is produced by an electronic data system and is valid without signature.


Traceable to one or more of the following National Metrology Institutes: NIM-China, NIST-USA, and VSL-The Netherlands

Micro Motion calibration facilities are accredited in accordance with the recognized International Standard ISO/IEC 17025. This Certificate is not covered by the available ISO/IEC 17025 accreditation.

28.0.0.301

2025.08.25 17:23:22

1 / 1



Flow (%)	Flow Rate (kg/min)	Meter Total (kg)	Reference Total (kg)	Error (%)	Specification (±%)
100.0	38.6	38.25003	38.2496	0.001	0.500
10.0	3.86	7.859089	7.857804	0.016	0.500
50.0	19.3	19.74454	19.74387	0.003	0.500
100.0	38.6	38.26611	38.26361	0.007	0.500

Micro Motion, Inc.

Transmitter Configuration Report

12294403

Product Code

CNG050S290NCAAEZZZ

2700I13ABAEZWZ

PUCK700

Serial ID

15623624

12294403

34943831

Order ID

5616237

5616237

Line

1.1

1.32

Item

2

2

Customer Tag

26011539

Process

Process ID : 1.38708762

Process Time : 2025.08.26 17:09:14

Process Stand : SSCB-CONFIG1@SSCB

Sensor

D1 : 0

D2 : 1

DFQ1 : 0

DFQ2 : 0

DT : 4.25

DTG : 0

Dens Press Comp Factor : 0

Density Meter Factor : 1

FCF : 141.52

FD : 0

FFQ : 0

FT : 4.5

FTG : 0

Flow Cal Pressure PSI : 30

Flow Press Comp Factor : 0

K1 : 4060.75

K2 : 4228.09

Mass Flow Meter Factor : 1

Volume Flow Meter Factor : 1

Special Mass Time Unit : SEC

Special Mass Total Text : NONE

Special Volume Base Unit : L

Special Volume Conv Factor : 1

Special Volume Flow Text : NONE

Special Volume Time Unit : SEC

Special Volume Total Text : NONE

Temperature Unit : C

Volume Flow Unit : L/MIN

MVD Channel Assignments

Channel A Assignment : Milliamp (Primary) Output [#1]

Channel B Power : Active (internally powered)

Assignments

Event 1 Variable : Density

Event 2 Variable : Density

Frequency1 Scaling Method : Frequency = Flow

Frequency1 Variable : Mass Flow Rate

mA1 Variable : Mass Flow Rate

Ranges

Event 1 Setpoint : 0

Event 1 Type : Event Low (Event "OFF" if PV > SP)

Event 2 Setpoint : 0

Event 2 Type : Event Low (Event "OFF" if PV > SP)

Frequency1 Active State : Active High

Frequency1 Hertz : 1000

Frequency1 Output Mode : Single

Frequency1 Pulses/Unit : 1.554404

Frequency1 Rate : 643.3333

Frequency1 Units/Pulse : 0.64333333

mA1 LRV : 0

mA1 URV : 643.3333

Units

Density Unit : G/CM3

GSV Flow Unit : SCFM

Mass Flow Unit : G/SEC

Pressure Unit : POUNDS/SQUARE INCH

Special GSV Base Time Unit : MIN

Special GSV Base Volume Unit : Standard cubic feet

Special GSV Conv Factor : 1

Special GSV Flow Unit Text : NONE

Special GSV Total Text : NONE

Special Mass Base Unit : G

Special Mass Conv Factor : 1

Special Mass Flow Text : NONE

Faults

28.0.0.311

2025.08.26 17:19:29

1/2

Faults

Frequency1 Fault Behavior : Upscale
Frequency1 Fault Value : 15000
mA1 Fault Behavior : Downscale (Default)
mA1 Fault Value : 2

Other

Calibration Process ID : 9.31175738
Core Software Rev : 35
Density Cutoff : 0.2
Density Damping : 1.6
Density High Limit : 5
Density Low Limit : 0
Direction : FORWARD
Fault Dwell Time : 0
Feature Key : 1
Flow Damping : 0.8
HART Device ID : 4951801
LD Type : 0
Mass Flow Cutoff : 1.836
Pressure Comp Line Pressure : 0
Pressure Compensation State : OFF
RS485 Baud : 1200 baud
RS485 Parity : Odd
RS485 Protocol : HART
Density Slug Duration : 0
Tag :
Temperature Damping : 4.8
Transmitter Software Rev : 80
Volume Flow Cutoff : 0.11016