

Product Code

CNG050S290NCAAEZZZ

Serial ID

15652187

Order ID

6491418

Line

1.1

Item

1

Customer Tag

SN 26071558

Process

Process ID : 9.31353307

Process Time : 2026.06.30 20:18:39

Process Stand : TSM2G@SSCCL:1

Stand Uncertainty : +/-0.030%

Fluid : H2O

100% Rate : 38.6 KG/MIN

Pickoff : 1

Max Rate P/T : 34.42 PSIG/26.8 C

Results

Status : PASS

D1 : 0

D2 : 1

K1 : 4072.915

K2 : 4241.399

DT : 4.25

FD : 0

DTG : 0

DFQ1 : 0

DFQ2 : 0

FlowCal : 140.854.50

FFQ : 0

FTG : 0

DensCal : 04073042414.25

FCF : 140.85

FT : 4.5

ROLAND, VARGA


Technician

This certificate is produced by an electronic data system and is valid without signature.

Traceable to one or more of the following National Metrology Institutes: NIM-China, NIST-USA, and VSL-The Netherlands
Micro Motion calibration facilities are accredited in accordance with the recognized International Standard ISO/IEC 17025. This Certificate is not covered by the available ISO/IEC 17025 accreditation.

28.0.0.317 2026.07.20 18:57:31

1 / 1



Flow (%)	Flow Rate (kg/min)	Meter		Reference		Error (%)	Specification (±%)
		Total (kg)		Total (kg)			
100.0	38.6	38.2957		38.29169		0.010	0.500
10.0	3.86	11.66453		11.66839		-0.033	0.500
50.0	19.3	19.82914		19.82692		0.011	0.500
100.0	38.6	38.24874		38.24435		0.011	0.500

Product Code	Serial ID	Order ID	Line	Item	Customer Tag
CNG050S290NCAAEZZZ	15652187	6491418	1.1	1	
2700I13ABAEZWZ	12316951	6491418	1.31	1	
PUCK800	26332881				

S/N 26071558

Process

Process ID : 1.39314755
Process Time : 2026.07.21 17:51:46
Process Stand : SSCB-CONFIG1@SSCB

Sensor

Units

D1 : 0
D2 : 1
DFQ1 : 0
DFQ2 : 0
DT : 4.25
DTG : 0
Dens Press Comp Factor : 0
Density Meter Factor : 1
FCF : 140.85
FD : 0
FFQ : 0
FT : 4.5
FTG : 0
Flow Cal Pressure PSI : 30
Flow Press Comp Factor : 0
K1 : 4072.915
K2 : 4241.399
Mass Flow Meter Factor : 1
Volume Flow Meter Factor : 1

Special Mass Time Unit : SEC
Special Mass Total Text : NONE
Special Volume Base Unit : L
Special Volume Conv Factor : 1
Special Volume Flow Text : NONE
Special Volume Time Unit : SEC
Special Volume Total Text : NONE
Temperature Unit : C
Volume Flow Unit : L/MIN

MVD Channel Assignments

Channel A Assignment : Millilamp (Primary) Output [#1]
Channel B Power : Active (internally powered)

Assignments

Event 1 Variable : Density
Event 2 Variable : Density
Frequency1 Scaling Method : Frequency = Flow
Frequency1 Variable : Mass Flow Rate
mA1 Variable : Mass Flow Rate

Ranges

Event 1 Setpoint : 0
Event 1 Type : Event Low (Event "OFF" if PV > SP)
Event 2 Setpoint : 0
Event 2 Type : Event Low (Event "OFF" if PV > SP)
Frequency1 Active State : Active High
Frequency1 Hertz : 1000
Frequency1 Output Mode : Single
Frequency1 Pulses/Unit : 1.554404
Frequency1 Rate : 643.3333
Frequency1 Units/Pulse : 0.6433333
mA1 LRV : 0
mA1 URV : 643.3333

Faults

Units

Density Unit : G/CM3
GSV Flow Unit : SCFM
Mass Flow Unit : G/SEC
Pressure Unit : POUNDS/SQUARE INCH
Special GSV Base Time Unit : MIN
Special GSV Base Volume Unit : Standard cubic feet
Special GSV Conv Factor : 1
Special GSV Flow Unit Text : NONE
Special GSV Total Text : NONE
Special Mass Base Unit : G
Special Mass Conv Factor : 1
Special Mass Flow Text : NONE

Faults

Frequency1 Fault Behavior : Upscale

Frequency1 Fault Value : 15000

mA1 Fault Behavior : Downscale (Default)

mA1 Fault Value : 2

Other

Calibration Process ID : 9.31353307

Core Software Rev : 57

Density Cutoff : 0.2

Density Damping : 1.6

Density High Limit : 5

Density Low Limit : 0

Direction : FORWARD

Fault Dwell Time : 0

Feature Key : 1

Flow Damping : 0.8

HART Device ID : 5120859

LD Type : 0

Mass Flow Cutoff : 1.836

Pressure Comp Line Pressure : 0

Pressure Compensation State : OFF

RS485 Baud : 1200 baud

RS485 Parity : Odd

RS485 Protocol : HART

Density Slug Duration : 0

Tag :

Temperature Damping : 4.8

Transmitter Software Rev : 80

Volume Flow Cutoff : 0.11016

3/N 2607/558



Tulsa Gas Technologies, Inc.
4809 S. 101st East Ave Tulsa, OK 74146
PHONE: 918-665-2641 FAX: 918-665-2657

8/4/2026

Dispenser Serial Number 26071558

Micro Motion Transmitter Configuration

Required settings for correct operation of Micro Motion mass flow meter.

Transmitter Model Number: 2700
Sensor Model Number: CNG050
Transmitter Serial Number: 12316951
Sensor Serial Number: 15652187
Flow Calibration Factor: 140.854.50
Flow Units: LB/min

Communication on RS-485

Protocol	Modbus ASCII 7 Bit
Modbus Address	1
Baud Rate	9600
Parity	Even
Stop Bits	1

HART Communication

Superimposed on Primary mA (PV)

Analog Output (4-20 mA)

Analog Variable (PV)	Mass Flow
Lower Range Value	0.2500 lb/min
Upper Range Value	300.000 lb/min
mA Cutoff	0.0000 lb/min

Freq/Rate

Frequency variable (TV)	Mass Flow
Frequency Cutoff	0.2500 lb/min
Pulses per Unit	1000.00000 per lb

Temperature

Temp Units	deg F
------------	-------