

Product Code

CNG050S290NCAAEZZZ

Serial ID

15098259

Order ID

10392206

Line Item

2.1 2

Customer Tag

2101306A

Process



Process ID : 9.27640921
Process Time : 2020.11.16 12:46:49

Process Stand : SSF1E@SSCCL:1

Stand Uncertainty : +/-0.030%

Fluid : H2O

100% Rate : 38.6 KG/MIN

Pickoff : 1

Max Rate P/T : 41.64 PSIG/23.6 C

Results

Status : PASS

D1 : 0

D2 : 1

K1 : 4032.522

K2 : 4198.846

DT : 4.25

FD : 0

DTG : 0

DFQ1 : 0

DFQ2 : 0

FlowCal : 142.284.50

FFQ : 0

FTG : 0

DensCal : 04033041994.25

FCF : 142.28


FT : 4.5

[Signature]

RAZVAN OCTAVIAN, COSMA

Technician

This certificate is produced by an electronic data system and is valid without signature.



Flow (%)	Flow Rate (kg/min)	Meter Total (kg)	Reference Total (kg)	Error (%)	Specification (±%)
100.0	38.6	38.33495	38.33684	-0.005	0.500
10.0	3.86	7.865007	7.865688	-0.009	0.500
50.0	19.3	19.75741	19.75655	0.004	0.500
100.0	38.6	38.26868	38.26874	0.000	0.500

2107306A

Product Code

CNG050S290NCAAEZZZ

2700I13ABAEZWZ

PUCK700

Serial ID

15098259

3480832

26112740

Order ID

10392206

10392206

Line Item

2.1 2

2.34 2

Customer Tag

Process

Process ID : 1.36090861

Process Time : 2021.01.26 18:47:10

Process Stand : CONFIGURATION@SSCB

Sensor

D1 : 0

D2 : 1

DFQ1 : 0

DFQ2 : 0

DT : 4.25

DTG : 0

Dens PCF : 0

Density Meter Factor : 1

FCF : 142.28

FD : 0

FFQ : 0

FT : 4.5

FTG : 0

Flow PCP : 30

Flow PCF : 0

K1 : 4032.522

K2 : 4198.846

Mass Flow Meter Factor : 1

Volume Flow Meter Factor : 1

Units

Special Mass Time Unit : SEC

Special Mass Total Text : NONE

Special Volume Base Unit : L

Special Volume Conv Factor : 1

Special Volume Flow Text : NONE

Special Volume Time Unit : SEC

Special Volume Total Text : NONE

Temperature Unit : C

Volume Flow Unit : L/MIN

MVD Channel Assignments

Channel B Power : Active (internally powered)

Assignments

Event 1 Variable : Density

Event 2 Variable : Density

Frequency1 Scaling Method : Frequency = Flow

Frequency Variable 1 : Mass Flow Rate

mA1 Variable : Mass Flow Rate

Ranges

Event 1 Setpoint : 0

Event 1 Type : Event Low (Event "OFF" if PV > SP)

Event 2 Setpoint : 0

Event 2 Type : Event Low (Event "OFF" if PV > SP)

Frequency1 Active State : Active High

Frequency1 Hertz : 1000

Frequency1 Output Mode : Single

Frequency1 Pulses/Unit : 1.554404

Frequency1 Rate : 643.3333

Frequency1 Units/Pulse : 0.64333333

mA1 LRV : 0

mA1 URV : 643.3333

Faults

Frequency1 Fault Behavior : Upscale

Units

Density Unit : G/CM3

GSV Flow Unit : SCFM

Mass Flow Unit : G/SEC

Pressure Unit : POUNDS/SQUARE INCH

Special GSV Base Time Unit : MIN

Special GSV Base Volume Unit : Standard cubic feet

Special GSV Conv Factor : 1

Special GSV Flow Unit Text : NONE

Special GSV Total Text : NONE

Special Mass Base Unit : G

Special Mass Conv Factor : 1

Special Mass Flow Text : NONE

Faults

Frequency1 Fault Value : 15000
mA1 Fault Behavior : Downscale (Default)
mA1 Fault Value : 2

Other

Calibration Process ID : 9.27640921
Core Software Rev : 35
Density Cutoff : 0.2
Density Damping : 1.6
Density High Limit : 5
Density Low Limit : 0
Direction : FORWARD
Fault Dwell Time : 0
Feature Key : 1
Flow Damping : 0.8
HART Device ID : 6120750
LD Type : 0
Mass Flow Cutoff : 1.836
Pressure Comp Line Pressure : 0
Pressure Compensation State : OFF
RS485 Baud : 1200 baud
RS485 Parity : Odd
RS485 Protocol : HART
Slug Duration : 0
Tag :
Temperature Damping : 4.8
Transmitter Software Rev : 80
Volume Flow Cutoff : 0.11016

21011306A



Tulsa Gas Technologies, Inc.
4809 S. 101st East Ave Tulsa, OK 74146
PHONE: 918-665-2641 FAX: 918-665-2657

2/2/2021

Dispenser Serial Number 21011306

Side A

Micro Motion Transmitter Configuration

Required settings for correct operation of Micro Motion mass flow meter.

Transmitter Model Number: 2700
Sensor Model Number: CNG095
Transmitter Serial Number: 3480832
Sensor Serial Number: 15098259
Flow Calibration Factor: 142.284.50
Flow Units: lb/min

Communication on RS-485

Protocol: Modbus ASCII 7 Bit
Modbus Address: 1
Baud Rate: 9600
Parity: Even
Stop Bits: 1

HART Communication

Superimposed on Primary mA (PV)

Analog Output (4-20 mA)

Analog Variable (PV): Mass Flow
Lower Range Value: 00000 lb/min
Upper Range Value: 300.000 lb/min
mA Cutoff: 0.0000 lb/min

Freq/Rate

Frequency variable (TV): Mass Flow
Frequency Cutoff: 0.2500 lb/min
Pulses per Unit: 1000.00000 per lb

Temperature

Temp Units: deg F

Product Code

CNG050S290NCAAEZZZ

Serial ID

15098182

Order ID

10392206

Line Item

1.1 2

Customer Tag

21011306#

Process

Process ID : 9.27628615

Process Time : 2020.11.13 9:31:09

Process Stand : SSF1E@SSCCL:1

Stand Uncertainty : +/-0.030%

Fluid : H2O

100% Rate : 38.6 KG/MIN

Pickoff : 1

Max Rate P/T : 41.7 PSIG/22.8 C

Results

Status : PASS

D1 : 0

D2 : 1

K1 : 4040.784

K2 : 4205.171

DT : 4.25

FD : 0

DTG : 0

DFQ1 : 0

DFQ2 : 0

FlowCal : 143.074.50

FFQ : 0

FTG : 0

DensCal : 04041042054.25

FCF : 143.07


FT : 4.5

[Signature]

BUNGARDEAN, OVIDIU

Technician

This certificate is produced by an electronic data system and is valid without signature.



Flow (%)	Flow Rate (kg/min)	Meter Total (kg)	Reference Total (kg)	Error (%)	Specification (±%)
100.0	38.6	38.30342	38.30315	0.001	0.500
10.0	3.86	7.885723	7.888806	-0.039	0.500
50.0	19.3	19.73843	19.73775	0.003	0.500
100.0	38.6	38.30921	38.31018	-0.003	0.500

Product Code	Serial ID	Order ID	Line	Item	Customer Tag
CNG050S290NCAFAEZZZ	15098182	10392206	1.1	2	
2700I13ABAEZHZ	3481951	10392206	1.34	2	
PUCK700	26119071				

Process

Process ID : 1.36090896
 Process Time : 2021.01.26 19:20:55
 Process Stand : CONFIGURATION@SSCB

Sensor

D1 : 0
 D2 : 1
 DFQ1 : 0
 DFQ2 : 0
 DT : 4.25
 DTG : 0
 Dens PCF : 0
 Density Meter Factor : 1
 FCF : 143.07
 FD : 0
 FFQ : 0
 FT : 4.5
 FTG : 0
 Flow PCP : 30
 Flow PCF : 0
 K1 : 4040.784
 K2 : 4205.171
 Mass Flow Meter Factor : 1
 Volume Flow Meter Factor : 1

Units

Special Mass Time Unit : SEC
 Special Mass Total Text : NONE
 Special Volume Base Unit : L
 Special Volume Conv Factor : 1
 Special Volume Flow Text : NONE
 Special Volume Time Unit : SEC
 Special Volume Total Text : NONE
 Temperature Unit : C
 Volume Flow Unit : L/MIN

MVD Channel Assignments

Channel B Power : Active (internally powered)

Assignments

Event 1 Variable : Density
 Event 2 Variable : Density
 Frequency1 Scaling Method : Frequency = Flow
 Frequency Variable 1 : Mass Flow Rate
 mA1 Variable : Mass Flow Rate

Ranges

Event 1 Setpoint : 0
 Event 1 Type : Event Low (Event "OFF" if PV > SP)
 Event 2 Setpoint : 0
 Event 2 Type : Event Low (Event "OFF" if PV > SP)
 Frequency1 Active State : Active High
 Frequency1 Hertz : 1000
 Frequency1 Output Mode : Single
 Frequency1 Pulses/Unit : 1.554404
 Frequency1 Rate : 643.3333
 Frequency1 Units/Pulse : 0.64333333
 mA1 LRV : 0
 mA1 URV : 643.3333

Faults

Frequency1 Fault Behavior : Upscale

Units

Density Unit : G/CM3
 GSV Flow Unit : SCFM
 Mass Flow Unit : G/SEC
 Pressure Unit : POUNDS/SQUARE INCH
 Special GSV Base Time Unit : MIN
 Special GSV Base Volume Unit : Standard cubic feet
 Special GSV Conv Factor : 1
 Special GSV Flow Unit Text : NONE
 Special GSV Total Text : NONE
 Special Mass Base Unit : G
 Special Mass Conv Factor : 1
 Special Mass Flow Text : NONE

Faults

Frequency1 Fault Value : 15000
mA1 Fault Behavior : Downscale (Default)
mA1 Fault Value : 2

Other

Calibration Process ID : 9.27628615
Core Software Rev : 35
Density Cutoff : 0.2
Density Damping : 1.6
Density High Limit : 5
Density Low Limit : 0
Direction : FORWARD
Fault Dwell Time : 0
Feature Key : 1
Flow Damping : 0.8
HART Device ID : 6120754
LD Type : 0
Mass Flow Cutoff : 1.836
Pressure Comp Line Pressure : 0
Pressure Compensation State : OFF
RS485 Baud : 1200 baud
RS485 Parity : Odd
RS485 Protocol : HART
Slug Duration : 0
Tag :
Temperature Damping : 4.8
Transmitter Software Rev : 80
Volume Flow Cutoff : 0.11016

2101306B



Tulsa Gas Technologies, Inc.
4809 S. 101st East Ave Tulsa, OK 74146
PHONE: 918-665-2641 FAX: 918-665-2657

2/2/2021

Dispenser Serial Number 21011306

Side B

Micro Motion Transmitter Configuration

Required settings for correct operation of Micro Motion mass flow meter.

Transmitter Model Number: 2700
Sensor Model Number: CNG095
Transmitter Serial Number: 3481951
Sensor Serial Number: 15098182
Flow Calibration Factor: 143.074.50
Flow Units: lb/min

Communication on RS-485

Protocol: Modbus ASCII 7 Bit
Modbus Address: 1
Baud Rate: 9600
Parity: Even
Stop Bits: 1

HART Communication

Superimposed on Primary mA (PV)

Analog Output (4-20 mA)

Analog Variable (PV): Mass Flow
Lower Range Value: 00000 lb/min
Upper Range Value: 300.000 lb/min
mA Cutoff: 0.0000 lb/min

Freq/Rate

Frequency variable (TV): Mass Flow
Frequency Cutoff: 0.2500 lb/min
Pulses per Unit: 1000.00000 per lb

Temperature

Temp Units: deg F