

Product Code

CNG050S290NCAAEZZZ

Serial ID

15331435

Order ID

10462906

Line

1.1

Item Customer Tag

22101341

Process

Process ID : 9.296663969

Process Time : 2022.03.09 18:10:32

Process Stand : SSF1E@SSCCL:1

Stand Uncertainty : +/-0.030%

Fluid : H2O

100% Rate : 38.6 KG/MIN

Pickoff : 1

Max Rate P/T : 41.82 PSIG/20.4 C

Results

Status : PASS

D1 : 0

D2 : 1

K1 : 4100.544

K2 : 4270.802

DT : 4.25

FD : 0

DTG : 0

DFQ1 : 0

DFQ2 : 0

FlowCal : 139.094.50

FFQ : 0



FTG : 0

DensCal : 04101042714.25

FCF : 139.09

FT : 4.5

Detail



Flow (%)	Flow Rate (kg/min)	Meter Total (kg)	Reference Total (kg)	Error (%)	Specification (±%)
100.0	38.6	38.28155	38.28352	-0.005	0.500
10.0	3.86	7.865651	7.870394	-0.060	0.500
50.0	19.3	19.80277	19.80168	0.006	0.500
100.0	38.6	38.27962	38.27959	0.000	0.500

By [Signature]
 RAZVAN, TATAR
 Technician

This certificate is produced by an electronic data system and is valid without signature.

Product Code

CNG050S290NCAAEZZZ

2700I13ABAEZMZ

PUC700

Serial ID

15331435

3520023

26196089

Order ID

10462906

10462906

Line Item

1.1 1

1.33 1

Customer Tag

22101341

Process

Process ID : 1.36699674

Process Time : 2022.04.04 17:54:05

Process Stand : CONFIGURATION@SSCB

Sensor

D1 : 0

D2 : 1

DFQ1 : 0

DFQ2 : 0

DT : 4.25

DTG : 0

Dens PCF : 0

Density Meter Factor : 1

FCF : 139.09

FD : 0

FFQ : 0

FT : 4.5

FTG : 0

Flow PCP : 30

Flow PCF : 0

K1 : 4100.544

K2 : 4270.802

Mass Flow Meter Factor : 1

Volume Flow Meter Factor : 1

Units

Special Mass Time Unit : SEC

Special Mass Total Text : NONE

Special Volume Base Unit : L

Special Volume Conv Factor : 1

Special Volume Flow Text : NONE

Special Volume Time Unit : SEC

Special Volume Total Text : NONE

Temperature Unit : C

Volume Flow Unit : L/MIN

MVD Channel Assignments

Channel B Power : Active (internally powered)

Assignments

Event 1 Variable : Density

Event 2 Variable : Density

Frequency1 Scaling Method : Frequency = Flow

Frequency Variable 1 : Mass Flow Rate

mA1 Variable : Mass Flow Rate

Ranges

Event 1 Setpoint : 0

Event 1 Type : Event Low (Event "OFF" if PV > SP)

Event 2 Setpoint : 0

Event 2 Type : Event Low (Event "OFF" if PV > SP)

Frequency1 Active State : Active High

Frequency1 Hertz : 1000

Frequency1 Output Mode : Single

Frequency1 Pulses/Unit : 1.554404

Frequency1 Rate : 643.3333

Frequency1 Units/Pulse : 0.64333333

mA1 LRV : 0

mA1 URV : 643.3333

Faults

Frequency1 Fault Behavior : Upscale

Units

Density Unit : G/CM3

GSV Flow Unit : SCFM

Mass Flow Unit : G/SEC

Pressure Unit : POUNDS/SQUARE INCH

Special GSV Base Time Unit : MIN

Special GSV Base Volume Unit : Standard cubic feet

Special GSV Conv Factor : 1

Special GSV Flow Unit Text : NONE

Special GSV Total Text : NONE

Special Mass Base Unit : G

Special Mass Conv Factor : 1

Special Mass Flow Text : NONE

Faults

Frequency1 Fault Value : 15000
mA1 Fault Behavior : Downscale (Default)
mA1 Fault Value : 2

Other

Calibration Process ID : 9.29663969

Core Software Rev : 35
Density Cutoff : 0.2
Density Damping : 1.6
Density High Limit : 5
Density Low Limit : 0
Direction : FORWARD
Fault Dwell Time : 0
Feature Key : 1
Flow Damping : 0.8
HART Device ID : 6193000
LD Type : 0
Mass Flow Cutoff : 1.836
Pressure Comp Line Pressure : 0
Pressure Compensation State : OFF
RS485 Baud : 1200 baud
RS485 Parity : Odd
RS485 Protocol : HART
Slug Duration : 0
Tag :
Temperature Damping : 4.8
Transmitter Software Rev : 80
Volume Flow Cutoff : 0.11016



Tulsa Gas Technologies, Inc.
4809 S. 101st East Ave Tulsa, OK 74146
PHONE: 918-665-2641 FAX: 918-665-2657

10/20/2022

Dispenser Serial Number 22101341

Micro Motion Transmitter Configuration

Required settings for correct operation of Micro Motion mass flow meter.

Transmitter Model Number: 2700
Sensor Model Number: CNG095
Transmitter Serial Number: 3520023
Sensor Serial Number 15331435
Flow Calibration Factor: 139.094.50
Flow Units lb/min

Communication on RS-485

Protocol Modbus ASCII 7 Bit
Modbus Address 1
Baud Rate 9600
Parity Even
Stop Bits 1

HART Communication

Superimposed on Primary mA (PV)

Analog Output (4-20 mA)

Analog Variable (PV) Mass Flow
Lower Range Value 00000 lb/min
Upper Range Value 300.000 lb/min
mA Cutoff 0.0000 lb/min

Freq/Rate

Frequency variable (TV) Mass Flow
Frequency Cutoff 0.2500 lb/min
Pulses per Unit 1000.00000 per lb

Temperature

Temp Units deg F