

Tulsa Gas Technologies, Inc.

Calibration Report

4809 South 101st East Ave.
Tulsa, OK 74146
Phone 918-665-2641 Fax 918-665-2657

DATE: February 20, 2019
Test #
Test Date: 2/20/2019
Tested By: Josh Breedlove/Dylan Higginbotham
Test Type: Acceptance

Test Site:
Tulsa Gas Technologies, Inc.
4809 South 101st East Ave.
Tulsa, OK 74146

CONTACT: Jerry Kelley
PHONE: 918 665 2641

Work Order Number:
TRIL120517

Customer Name:
Tulsa Gas Technologies
4809 South 101st East Ave.
Tulsa, OK. 74146

Dispenser Model Number:
T-7214-12-8CNG50

Dispenser Serial Number:
18011204

Eprom Number:
1.07.02

	SCALE	DISPENSER	PERCENT of ERROR
HOSE 1	MASS/GGE	GGE	
	10.38	10.410	-0.29%
	10.10	10.090	0.10%
	9.42	9.380	0.42%
HOSE 2	8.62	8.510	1.28%
	10.20	10.290	-0.88%
	9.30	9.280	0.22%

EXPECTABLE TOLERANCE IS + - 2% FOR EACH BATCH
AND + - .8% BETWEEN BATCHES PER EACH HOSE