

Tulsa Gas Technologies, Inc.

Calibration Report

4809 South 101st East Ave.
Tulsa, OK 74146
Phone 918-665-2641 Fax 918-665-2657

DATE: August 28, 2018
Test #
Test Date: 8/28/2018
Tested By: Josh Breedlove/C.R. Kendal
Test Type Acceptance

Test Site:
Tulsa Gas Technologies, Inc.
4809 South 101st East Ave.
Tulsa, OK 74146

CONTACT: Jerry Kelley
PHONE: 918 665 2641

Work Order Number:
PennDOT-Butler

Customer Name:
Tulsa Gas Technologies
4809 South 101st East Ave.
Tulsa, OK. 74146

Dispenser Model Number:
T-7204-12CNG50W

Dispenser Serial Number:
18041214

Eprom Number:
1.04.21

	SCALE	DISPENSER	PERCENT of ERROR
HOSE 1	MASS/GGE	GGE	
	9.92	9.970	-0.50%
	10.07	10.100	-0.30%
	9.91	9.890	0.20%
HOSE 2	10.60	10.630	-0.28%
	10.60	10.610	-0.09%
	10.28	10.300	-0.19%

EXPECTABLE TOLERANCE IS + - 2% FOR EACH BATCH
AND + - .8% BETWEEN BATCHES PER EACH HOSE