

# Tulsa Gas Technologies, Inc.

# Calibration Report

4809 South 101st East Ave.  
Tulsa, OK 74146  
Phone 918-665-2641 Fax 918-665-2657

**DATE:** August 28, 2018  
**Test #**  
**Test Date:** 8/28/2018  
**Tested By:** Josh Breedlove/C.R. Kendal  
**Test Type** Acceptance

**Test Site:**  
Tulsa Gas Technologies, Inc.  
4809 South 101st East Ave.  
Tulsa, OK 74146

**CONTACT:** Jerry Kelley  
**PHONE:** 918 665 2641

**Work Order Number:**  
PennDOT-Erie

**Customer Name:**  
Tulsa Gas Technologies  
4809 South 101st East Ave.  
Tulsa, OK. 74146

**Dispenser Model Number:**  
T-7104-12CNG50W

**Dispenser Serial Number:**  
18041212

**Eprom Number:**  
1.04.70

	SCALE	DISPENSER	PERCENT of ERROR
<b>HOSE 1</b>	<b>MASS/GGE</b>	<b>GGE</b>	
	9.92	9.970	-0.50%
	10.07	10.100	-0.30%
	9.91	9.890	0.20%
<b>HOSE 2</b>	1.00	1.000	0.00%
	1.00	1.000	0.00%
	1.00	1.000	0.00%

EXPECTABLE TOLERANCE IS + - 2% FOR EACH BATCH  
AND + - .8% BETWEEN BATCHES PER EACH HOSE