

# Tulsa Gas Technologies, Inc.

4809 South 101st East Ave.  
Tulsa, OK 74146  
Phone 918-665-2641 Fax 918-665-2657

# Calibration Report

**DATE:** December 28, 2016

**Test #**

**Test Date:** 12/28/2016

**Tested By:** Dylan Higginbotham/J.T. Breedlove

**Test Type:** Acceptance

**Test Site:**

Tulsa Gas Technologies, Inc.  
4809 South 101st East Ave.  
Tulsa, OK 74146

**CONTACT:** Jerry Kelley

**PHONE:** 918 665 2641

**Work Order Number:**

TRIL01416A

**Customer Name:**

Tulsa Gas Technologies  
4809 South 101st East Ave.  
Tulsa, OK. 74146

**Dispenser Model Number:**

T7104-12CNG50W

**Dispenser Serial Number:**

16121142

**Eprom Number:**

1.04.70

	SCALE	DISPENSER	PERCENT of ERROR
HOSE 1	MASS/GGE	GGE	
	0.00	0.000	#DIV/0!
	0.00	0.000	#DIV/0!
	0.00	0.000	#DIV/0!
HOSE 2	6.43	6.460	-0.50%
	6.37	6.412	-0.61%
	7.42	7.454	-0.53%

EXPECTABLE TOLERANCE IS + - 2% FOR EACH BATCH  
AND + - .8% BETWEEN BATCHES PER EACH HOSE

